

UNIVERSITY OF CALICUT
(Purchase Division)

109527/PURCHASE-ASST-B2/2017/Admn

05.08.2017

TENDER NOTICE

Tender No.	109527/Purchase - Asst B2/2017/Admn Dated 05.08.2017
Name of equipment	PCR Machine
Department to which the Purchases is made	Department of Bio Technology
Quantity required	1 (one)
Last date & time of the receipt of the Tender	26.08.2017
Date & Time of opening the Tender	30.08.2017
Designation and Address of the officer to whom the Quotations are to be sent.	The Deputy Registrar, Purchase Division, University of Calicut. PIN.673 635. Phone.0494- 2407130

- Sealed competitive Tenders (in Two cover System) are invited for the purchase of ONE PCR Machine to the Department of Bio Technology.
- The Tenders should reach the Deputy Registrar (Purchase) on or before 26.08.2017. The Tenders received late, will be rejected summarily.
- The rate quoted should include base price, GST and other charges if any, separately and the consolidated amount as well.
- The equipment shall be delivered within ten days after receiving the supply order.
- GST registration is mandatory for the firms to participate in the tender and the copy of the GST registration Certificate should be attached with the tender documents.

The specifications and terms & conditions are as follows:-

Sealed superscribed tenders are invited for the purchase of PCR Machine (as per the specification noted below) to the Department of Bio Technology . The Tender form can be had from the Purchase Division or can be downloaded from the University Website. Sealed tenders with cost of tender form (0.2% of the cost of tender rounded to the nearest multiple of 100, subject to a minimum of Rs.400+GST @5%. and maximum Rs.1500+GST @5%.) and EMD @1% of the total cost of the articles tendered for subject to a minimum of Rs.1500/-) drawn in favour of the Finance Officer should be submitted to the Deputy Registrar(Purchase Division) University of Calicut.

A Performance Security equivalent to 5% of the total value of the contract rounded to the nearest rupee should be submitted by the successful bidder for a contract value above Rs.1,00,000/-. The EMD and Performance Security are returnable.

specifications

Sl No	Name of Equioment	Specification	Quandity required
1	PCR Machine	With Peltier technology Capacity - 96 X0.2 ml tubes Power 220-240 V AC Temp range - 4-99°C Ramp rate - 3°C /4 °c per sec Accuracy - ± 0.25 °c / ± 0.5 ° C With heated lid, use of flat capped tubes, USBport Preferably with auto-restart function Suitable stabilizer and online UPS-2 hours backup 2-3 years warranty	1

To the System Analyst, to publish in the University Web Site


Shajee K

Deputy Registrar


23/11