

**University Of Calicut  
(Purchase Division)**

13146/PURCHASE-ASST-B2/2017/Admn (I)

10.08.2017

QUOTATION NOTICE

Quotation No.	13146/Purchase/Asst -B2/2017 Admn Dated 07/08/2017
Name of equipment	Bench top pH Meter
Department to which the Purchases is made	Zoology
Quantity required	1 (one)
Last date & time of the receipt of the Quotation	29.08.2017
Date & Time of opening the Quotation	31.08.2017
Designation and Address of the officer to whom the Quotations are to be sent.	The Deputy Registrar, Purchase Division, University of Calicut. PIN.673 635. Phone.0494- 2407130

- Sealed competitive quotations are invited for the purchase of ONE Bench top pH Meter to the Department of Zoology
- The quotations should reach the Deputy Registrar (Purchase) on or before 29.08.2017. The quotations received late, will be rejected summarily.
- Rate quoted should include base price, GST and other charges if any, separately and the consolidated amount as well.
- The equipment shall be delivered within ten days after receiving the supply order.

The specifications and terms & conditions are as follows:-

Sl No	Name equipment	Specification	Nos required
1	Bench top pH Meter	<p>pH Range : 0.00 to 14.00 pH                      Resolution/Accuracy 0.01 pH / <math>\pm 0.01</math>pH                      No. of pH Calibration: Up to 3 points                      Points:USA- 4.01,7.00,10.01                      pH Buffer Options NIST- 4.01,6.86,9.18                      Temperature Compensation Automatic (optional) / Manual (0 to 100 °C)                      Temperature Range 0 to 100</p> <p>Specifications                      pH Range:0.00 to 14.00 pH                      Resolution/Accuracy:0.01 pH / <math>\pm 0.01</math>pH                      No. of pH Calibration Points:Up to 3 points                      pH Buffer Options USA- 4.01,7.00,10.01                      NIST- 4.01,6.86,9.18                      Temperature Compensation Manual (0 to 100)</p>	1

Terms & Conditions

- 1) The bidder should have GST registration and the copy of the certificate shall invariably be attached with the quotation documents.
- 2) The bidder should clearly mention the make and model of the items quoted.
- 3) The bidder must attach the technical data sheet of the items being quoted in the tender.

4) A comprehensive On-site warranty of minimum 1 year should be provided for the product.

Sd/-  
Deputy Registrar  
Purchase Division

To The System Analyst (To be published in the University site)