

**UNIVERSITY OF CALICUT
DEPARTMENT OF BOTANY**

216302/BOTSO/2017/Admn [2]

Calicut University PO dt 05.03.2018

Quotation Notice

Sealed quotations are invited by the Head, Department of Botany, University of Calicut for the purchase of a **pH Meter** for the Department of Botany. Details of quotations ,specifications of equipment, terms and conditions, General conditions are listed below.

Tender No	216302/BOT-SO/2017/Admn [2] dt 05.03.2018
Department	Department of Botany
Name of equipment	pH Meter
Last date and time for the receipt of the tender	20.03.2018 4 pm
Date and time for opening the tender	24.03.2018 11 am
Designation and address of officer to whom the tender is to be addressed	Head, Department of Botany, University of Calicut, Calicut university PO Malappuram- Malappuram-673635 Contact 673635 Contact No 0494-2407407 No 0494-2407407
Date upto which rates are to be firm	maximum 6 months
Superscription	Department of Botany, No216302/BOT-SO/2017/Admn [2] dt 05.03.2018

Specifications:-

1. pH METER WITH INDIGENOUS pH ELECTRODE

Specifications	
pH Range	0.00 to 14.00 pH
Resolution/Accuracy	0.01 pH / ± 0.01 pH
No. of pH Calibration Points	Up to 3 points

pH Buffer Options	USA- 4.01, 7.00, 10.01 NIST- 4.01,6.86,9.18
Temperature Compensation	Automatic (optional) / Manual (0 to 100 °C)
Temperature Range	0 to 100 °C
Temp Resolution / Accuracy	0.1 °C / ± 0.3 °C
pH Slope and Offset Display	Yes
Electrode	Included general purpose glass-body indigenous pH combination electrode AN-786-1, BNC connector, 1m cable length
Temperature Probe	To be ordered separately
Electrode Stand	Included Indian make
Power Requirement	AC/DC 9V Adapter (220 VAC, 500 mA)

2. Terms and conditions:-

- 1) The bidder should mention the make and model number.
- 2) The bidder should attach the technical data sheet of the product being quoted.
- 3 The bidder should have local presence with service centre in Kerala.

General conditions:-

1. The tender form can be downloaded from the university website [www.universityofcalicut.info] .
2. The bidder should have GST registration and the copy of the certificate in this regard shall invariably be attached with the tender documents. The rate quoted should be inclusive of all charges including taxes, installation charges, freight and delivery charges, etc if any. The base rate, GST and other charges if any should be shown separately in the tender.
3. The equipment shall be delivered within 10 days after receiving the supply order.
- 4 The quotations received late will be rejected without further notice. The firms which has been blacklisted already are not eligible to apply.
- 5 All Kerala Government rules related will be applicable to the quotations