

UNIVERSITY OF CALICUT
DEPARTMENT OF BOTANY

216302/BOTSO/2017/Admn

Calicut University PO dt 13.03.2018

Quotation Notice

Sealed quotations are invited by the Head, Department of Botany, University of Calicut for the purchase of a **Weighing Balance [0.0001 gm]** for the Department of Botany. Details of quotations

,specifications of equipment, terms and conditions, General conditions are listed below.

| | |
|--|--|
| Tender No | 216302/BOT-SO/2017/Admn dt 13.03.2018 |
| Department | Department of Botany |
| Name of equipment | Weighing Balance [0.0001gm] |
| Last date and time for the receipt of the tender | 28.03.2018 4 pm |
| Date and time for opening the tender | 02.04.2018 11 am |
| Designation and address of officer to whom the tender is to be addressed | Head, Department of Botany, University of Calicut, Calicut university PO Malappuram-673635 Contact No 0494-2407407 |
| Date upto which rates are to be firm | maximum 6 months |
| Superscription | Department of Botany, No216302/BOT-SO/2017/Admn dt 13 03.2018 |

Specifications:-

1. ELECTRONIC ANALYTICAL BALANCE

Specifications

- Capacity : 220 gms
- Readability : 0.0001g
- Volts : 12 V DC

- Repeatability Std. Deviation : ≤ 0.1 mg
- Linearity : ± 0.2 mg
- Response Time : 3secs.
- Operating ambient Temperature : 5 to 40° C
- Temp. coefficient of sensitivity 10-30° C : ± 2 ppm/ °C (When PSC is OFF)
- Sensitivity stability against temperature 10-30°C : ± 2 ppm (When PSC is ON)
- Pan Size : 80 mm dia approx.
- **Calibration : Internal**
- Body Dimensions : 220mmW x 330mmD x 310 mm H approx.
- Weight : 7kg approx.
- Power Consumption : 7 VA approx

Features

- PSC
- Built-in Clock
- GLP/ GMP/ ISO Calibration Report
- Window Direct
- Interval Timer Output
- RS-232 I/ F
- Specific Gravity measurement software
- Piece Counting
- % Display
- Unit Conversion
- Analog Display¹

1]The bidder should mention the make and model number.

2) The bidder should attach the technical data sheet of the product being quoted.

3 The bidder should have local presence with service centre in Kerala.

General conditions:-

1. The tender form can be downloaded from the university website [www.universityofcalicut.info] .
2. The bidder should have GST registration and the copy of the certificate in this regard shall invariably be attached with the tender documents. The rate quoted should be inclusive of all charges including taxes, installation charges, freight and delivery charges, etc if any. The base rate, GST and other charges if any should be shown separately in the tender.
3. The equipment shall be delivered within 10 days after receiving the supply order.
- 4 The quotations received late will be rejected without further notice. The firms which has been blacklisted already are not eligible to apply.
- 5 All Kerala Government rules related will be applicable to the quotations

